



Republic of the Philippines
PROVINCE OF BENGUET
 La Trinidad



BIDS AND AWARDS COMMITTEE (BeCI)
(DRUGS AND MEDICINES / MEDICAL SUPPLIES)

Bid Solicitation No.: SVP-2022-0052 March 3, 2022
 Purchase Request No.: 2022-03-0507 Date

MEMORANDUM TO:

Mr./Ms.: SALINIA L. CARBONEL
 (Canvasser)

Kindly provide appropriate dealers/ establishments copy of the attached request for quotation and/or solicitation to Bid for the Purchase of Newborn Hearing Screening Machine with an ABC of 500,000.00 to be used by ADH Their quotation /Bids shall be opened on MARCH 9, 2022

NAME OF ESTABLISHMENT	ADDRESS	RECEIVED		
		Name	Signature	Date
1.				
2.				
3.				
4.				
5.				
6.				
7.				
8.				
9.				
10.				

ATTY. BRIAN J. CRISPIN
Chairperson

CERTIFICATION

The undersigned, hereby jointly and severally certify to have given each of the above dealers / establishments a copy of the solicitation / bid papers and that no one, only those indicated above, failed to submit their quotations before the time of opening the same.

CANVASSER/S:

SALINIA L. CARBONEL
Signature over printed name/s



Purchase Request Number: **2022-03-0507**

Name of the Project: **Purchase of Newborn Hearing Screening Machine**

Location of the Project: **ADH**

Date: **March 3, 2022**

Quotation No.: **SVP-2022-0052**

CAFOA No.: **100220201992**

REQUEST FOR QUOTATION

Sir/Madam:

Kindly indicate your price quotation/s for the item/s listed below taking into consideration the specifications; submit this document duly signed by you or your authorized representative, to this office not late **9:00 am** on **WEDNESDAY** **MARCH 9, 2022** at the **BAC HEALTH, 1st Floor, Benguet General Hospital, La Trinidad, Benguet**.

Indicate also the shortest delivery period should you opt to. You may fax your quotations through Telefax No. 074-422-55-06 or thru e-mail at begh.bac@benguet.gov.ph

We may purchase the item/s from your company if the found to be reasonable and advantageous to the Benguet General Hospital, but we reserve the right to accept or reject the offer or all quotations not found in order.

Thank you very much.


ATTY. BRIAN J. CRISPIN
Chairperson

INSTRUCTIONS:

"PLEASE SEAL YOUR QUOTATIONS"

1. All entries must be Typewritten or handwritten (Please write legible and no use of pencil).
2. Delivery period within 14 (supplies, drugs and medicines, etc.) / 45 (Equipment) Calendar Days.
3. Warranty shall be for a period of **three (3) months** for supplies and materials and **one (1) year** for equipment reckoned from the date of acceptance by procuring entity.
4. Price validity shall be for a period of **120 calendar days**.
5. **Photocopy (Stamp "CERTIFIED TRUE COPY OF THE ORIGINAL" with name and signature)** of the following requirements to be submitted after opening and upon notification from the BAC.
 - a. Mayor's/ Business Permit.
 - b. PhilGEPS Registration No.
 - c. Submit original copy of complete (paragraph 1 to 10) Omnibus Sworn Statement (in addition to requirements a and b if the awarded item/s is above Php 50, 000.00).
 - d. Income / Business Tax Return (in addition to requirements a, b, and c if the awarded item/s is above PhP 500, 000).
 - e. Valid and current Certificate of Product Registration (CPR) issued by FDA of the DOH: if applicable
 - f. Brochure for Medical equipment: if applicable
 - g. Material Data Safety Sheet for Laboratories, Reagennts: if applicable
6. The procuring entity reserves the right to waive any defects in the tender or offer as the right to accept the bid most advantageous to the provincial
7. **ALL ERASURES SHOULD BE COUNTERSIGNED.**

Item No.	QTY	UNIT of ISSUE	ABC	ITEM & DESCRIPTION (SPECIFICATIONS)	BRAND/OFFER/ COMPLIANCE TO SPECIFICATIONS	UNIT PRICE	PRICE VALIDITY	DELIVERY PERIOD	TOTAL AMOUNT
			UNIT PRICE						
1	1		500,000.00	Newborn Hearing Screening Machine					
				<i>Measurement Types</i>					
				Screening and Diagnostic Testing DPOAE: 1.5 to 12 kHz, 40 to 70 dB SPL					
				TEOAE: 0.7 to 4 kHz, 83 dB / SPL					
				Handled Unit					
				Display Color OLED display					
				User Input: 4 - button Operation					
				Connectors:					
				Micro - USB for charging and communication					
				HDMI for probe					
				Weight: 6.4 oz (180 gm)					
				Communication to PC: Micro - USB					
				Languages: English, German, Spangish, French, Polish, Russian, Italian, Turkish, Portuguese, Chinese					
				Power Supply: 5.0v DC, 1.6A					
				<i>Probe</i>					
				Connector: HDMI					
				Probe Description:					
				Integrated microphone and receivers in probe head					
				Calibration data stored on probe					
				Cable Length: 40 in (101.6 cm)					
				Weight: 1.0 oz (28 gm)					
				Microphone noise: -20 dB SPL @ 2 kHz (1Hz)					
				Ear Tips: Single use disposable ear tips					
				<i>Cradle (Optional)</i>					
				Operation: Provides PC Database communication and charging					
				<i>Data</i>					
				Test Memory: 250 test on unit					
				Patient Names: Patient names on unit (optional)					
				Database Software: Report output to PDF, RTF, Image Files					
				<i>Printer (Optional)</i>					
				Type: Thermal Dot Matrix					
				Power: 7.4 V lithium ion battery 100 - 240 V, 50 / 60 Hz					
				Paper Width: 2.25 in (57 mm)					



Republic of the Philippines
 PROVINCE OF BENGUET
 La Trinidad
BIDS AND AWARDS COMMITTEE (BECH)
(DRUGS AND MEDICINES / MEDICAL SUPPLIES)



Purchase Request Number: **2022-03-0507**

Date: **March 3, 2022**

Name of the Project: **Purchase of Newborn Hearing Screening Machine**

Quotation No.: **SVP-2022-0052**

Location of the Project: **ADH**

CAFOA No.: **100220201992**

			Communication: Bluetooth				
			<i>Power</i>				
			Battery: 3.6 V rechargeable lithium ion				
			Battery Life: 15 hours on time				
			Charge Time: 4 hours to 100%				
			<i>Accessories</i>				
			Standard: Hand - held unit, Probe, Micro - USB charging cable for charger, Data software and micro - USB connector, Disposable ear tip kit and tubes, User Manual, Quick Guide, Calibration Certificate				
			Optional: Cradle, Printer, Carry Case, Ear Tips, Replacement Cables, Replacement Probe and Probe tubes				
			<i>General Standards</i>				
			IEC / EN 60601 - 1 Medical Electrical Equipment - Part 1: General requirements for basic safety and essential performance - 3rd edition				
			IEC / EN 60601 - 2 Medical Electrical Equipment - Part 1 - 2: General requirements for basic safety and essential performance - Collateral Standard: Electromagnetic Compatibility				
			UL 6061 - 1: Medical Electrical Equipment, Part1: General Requirement for Safety				
			CSA C22.2 # 601 - 1 - M90: Medical Electrical Equipment, Part 1: General Requirements for Safety				
			IEC 60645 - 6: Electronics - Audiometric Equipment - Part 6: Instruments for the measurement of otoacoustic emissions				
			ADH				

PURPOSE: To be use at Atok District Hospital **500,000.00**

CANVASSERS:

<p style="text-align: center;">CERTIFICATION</p> <p>I hereby certify that I have personally conducted this canvass and that the price/s quoted is/are true and correct and the signature of the Representative of the establishment who submitted the quotation/s is/are genuine.</p>	<p>After having carefully read and accepted your terms and conditions, I hereby submit the above quotations for your appropriate action.</p>
<p style="text-align: center;">SALINIA L. CARBONEL <hr/> PRINTED NAME & SIGNATURE OF AUTHORIZED CANVASSER</p>	<p style="text-align: center;">_____ (Name of Establishment per O.R. & address)</p> <p style="text-align: center;">_____ (Signature over Printed Name of Owner/Representative)</p> <p style="text-align: center;">_____ (Telephone/Cellphone Number or e-mail address)</p> <p style="text-align: center;">_____ Date</p>